



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/593,600	05/16/2007	Alfred Niederberger	10139/08802	1693

76960 7590 07/20/2010
Fay Kaplun & Marcin, LLP
150 Broadway, suite 702
New York, NY 10038

EXAMINER

LAWSON, MATTHEW JAMES

ART UNIT	PAPER NUMBER
----------	--------------

3775

MAIL DATE	DELIVERY MODE
-----------	---------------

07/20/2010

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/593,600	Applicant(s) NIEDERBERGER ET AL.	
	Examiner MATTHEW LAWSON	Art Unit 3775	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 01 June 2010.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 11, 12, 14 and 16-30 is/are pending in the application.
- 4a) Of the above claim(s) 22-29 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 11-12, 14, 16-21, 30 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>5/3/2010</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on June 1st, 2010 has been entered.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 11-12, 14, 16-19, and 21, are rejected under 35 U.S.C. 103(a) as being unpatentable over Macchia (US 3,054,321) in view of Landry (US 2004/0143265) in further view of Takachi (US 6,470,767).

Regarding claims 11, 14, 16, 18, and 21 Macchia discloses a bone screw comprising; a head (26), and a shaft extending distally from the head (figures 1-2, 10) and having a longitudinal axis, the shaft further comprising a proximal section (see figure below) adjoining the head and having a second outer diameter smaller than the

Art Unit: 3775

first outer diameter (figures 1-2, 10), an outer wall of the proximal section being at least partially threaded (34, figure 1), and a distal section (see figure below) extending distally from the proximal section and being attached to the proximal section by way of a ball (40)-and-socket (30a) joint (figures 1-2, 10), wherein the distal section is at least partially threaded (figures 1-2, 10) and the length of the shaft is constant (figures 1-2, 10), and the distal section is allowed to deflect no more than about 90 degrees relative to the longitudinal axis (figures 1-2, 10) and the distal section is rotatable relative to the proximal section about the longitudinal axis (column 2, lines 66-67).

Macchia fails to disclose the shaft being cannulated.

Landry et al. disclose a cannulated shaft (114, figure 6) to be used with a guide wire so that the shaft may be inserted into a vertebra at a desired location and in a desired angular orientation relative to the vertebra with limited or no visibility of the vertebrae (¶126). Accordingly it would have been obvious to one of ordinary skill in the art at the time the invention was made to have constructed the screw of Macchia to be cannulated so that a guide wire could be used permitting for insertion and alignment with limited or no visibility of the surgical site as taught by Landry.

Macchia further fails to disclose the head being spherical or cylindrical (having a first diameter) rather Macchia teaches a hexagonal head.

Macchia discloses the claimed invention except that the head is hexagonal instead of spherical. Takachi (US 6,470,767) shows that a hexagonal head is an equivalent structure known in the art (column 3, lines 34-37). Therefore, because these

Art Unit: 3775

two screw heads were art-recognized equivalents at the time the invention was made, one of ordinary skill in the art would have found it obvious to substitute the hexagonal head for the spherical head of Takachi.

Regarding claim 12, Macchia discloses the claimed invention except for the distal section comprising the socket portion of the connection and the proximal section comprising the ball portion of the ball-and-socket connection. It would have been obvious to one having ordinary skill in the art at the time the invention was made to have the ball portion be on the distal section and the socket portion of the proximal section since it has been held that a mere reversal of the essential working parts of a device involves only routine skill in the art. In re Einstein, 8 USPQ 167.

Regarding claim 17, Macchia discloses the claimed invention except for the proximal section having a larger diameter than the distal section. It would have been an obvious matter of design choice to have the distal section smaller than the proximal section, since applicant has not disclosed that the difference in diameters between the two sections solves any stated problem or is for any particular purpose and it appears that the invention would perform equally well with a proximal diameter that is equal or slightly smaller than the distal section.

With regard to claim 19, it would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the bone screw of the

Art Unit: 3775

combination of Macchia as modified by Landry and Takachi to deflect no more than about 30 degrees relative to the longitudinal axis, since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. In re Aller, 105 USPQ 233.

Claim 30 is rejected under 35 U.S.C. 103(a) as being unpatentable over Maccia (US 3,054,321) in view of Landry (US 2004/0143265).

Macchia discloses a bone screw comprising; a head (26), a shaft extending distally from the head and having a longitudinal axis, a proximal section (see figure below) adjoining the head wherein the proximal section is at least partially threaded (34), and a distal section (see figure below) polyaxially associated with the proximal section (figures 1-2, 8, and 10), the distal section being separated from the head by the proximal section (figures 1-2).

Macchia fails to disclose the shaft being cannulated.


Landry et al. disclose a cannulated shaft (114, figure 6) to be used with a guide wire so that the shaft may be inserted into a vertebra at a desired location and in a desired angular orientation relative to the vertebra with limited or no visibility of the vertebrae (¶126). Accordingly it would have been obvious to one of ordinary skill in the art at the time the invention was made to have constructed the screw of Macchia to be cannulated so that a guide wire could be used permitting for insertion and alignment with limited or no visibility of the surgical site as taught by Landry.

Art Unit: 3775

Claim 20 is rejected under 35 U.S.C. 103(a) as being unpatentable over Macchia (US 3,054,321) in view of Landry (US 2004/0143265) in view of Takachi (US 6,470,767) in further view of Purcell et al. (US 2004/0236330).

Macchia in view of Landry in view of Takachi disclose the claimed invention except for the ball having an octagonal shape.

Purcell et al. disclose a bone screw comprising a head, a shaft having a longitudinal axis, a proximal section adjoining the head, and a distal section; wherein the distal section is attached to the proximal section by way of a ball-and-socket joint (abstract), wherein the ball has a substantially octagonal shape (figures 12-13, ¶39) to permit a variable angular relationship between the two portions (¶39). Accordingly it would have been obvious to one of ordinary skill in the art at the time the invention was made to have constructed the device of Macchia to have a ball with an octagonal shape as taught by Purcell et al. to permit a variable angle relationship between the two portions of the bone screw.



Response to Arguments

Applicant's arguments with respect to all pending claims have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MATTHEW LAWSON whose telephone number is (571)270-7375. The examiner can normally be reached on M-F, 8am-5pm.

Art Unit: 3775

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Thomas Barrett can be reached on 571-272-4746. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/M. L./
Examiner, Art Unit 3775

/Thomas C. Barrett/
Supervisory Patent Examiner, Art
Unit 3775